■ *INTENDED USE*

The PPR Ab Test is a rapid test for the detection of PPR virus. It is a lateral flow immunochromatographic assay for the qualitative diagnosis of Peste des Petits Ruminants Virus Antibody (PPR Ab) in sheep's serum or plasma specimen.

Assay Time: 5-10 minutes

Specimen: serum, or plasma

■ *PRINCIPLE*

The PPR Ab Test is based on sandwich lateral flow immunochromatographic assay. The test device has a testing window for the observation of assay running and result reading. The testing window has an invisible T (test) zone and a C (control) zone before running the assay. When the treated sample was applied into the sample well on the device, the liquid will laterally flow through the surface of the test strip and react with the pre-coated PPR antigens. If there are anti-PPR antibodies in the specimen, a visible T line will appears. The C line should always appear after a sample is applied, which indicates a valid result. By this means, the device can accurately indicate the presence of Ovine Peste des Petits Ruminants Antibody in the specimen.

■ *REAGENTS AND MATERIALS*


- 10 test pouches, with cards and disposable droppers
- 1 package insert

■ *STORAGE AND STABILITY*

The kit can be stored at room temperature (2-30°C). The test kit is stable through the expiration date (18 months) marked on the package label. DO NOT FREEZE. Do not store the test kit in direct sunlight.

■ *TEST PROCEDURE*

- Collect sheep's fresh whole blood, get the serum or plasma specimen for use. Immediately apply the testing once collected the specimen. If not use right now, keep storage at 2-8°C. If the storage time more than 24hours, the specimen should be frozen at -20°C.
- Take the test kit to room temperature (20-30°C) before use. Keep all component in room temperature for at least 30 minutes to recover the reaction conditions.
- Take out the test device from the foil pouch and place it horizontally.
- Place 3 drops (approx. 120μL) of serum or plasma specimen into the sample well of the test card.
- Interpret the result at 5-10 minutes. Result after 15 minutes is considered as invalid.



■ *INTERPRETATION OF RESULTS*

- Positive (+): The presence of both "C" line and zone "T" line, no matter T line is clear or vague.
- Negative (-): Only clear C line appear. No T line.
- Invalid: No colored line appears in C zone. No matter if T line appears.

■ *PRECAUTIONS*

- Do not use tap water or other liquid as negative control. Only serum or plasma could be used in the assay.
- All reagents must be at room temperature before running the assay.
- Do not remove test cassette from its pouch until immediately before use.
- Do not use the test beyond its expiration date.
- The components in this kit have been quality control tested as standard batch unit. Do not mix components from different lot numbers.
- All specimens are of potential infection. It must be strictly treated according to the rules and regulations by local states.

■ *LIMITATION*

PPR Ab Test is for in vitro veterinary diagnosis use only. All result should be considered with other clinical information available with veterinarian. It is suggested to apply a further confirmative method when positive result was observed.

Manufacturer:

Hangzhou Evegen Biotech Co., LTD.

1st Floor, Building#1 No.578, 20th Street, Baiyang Subdistrict, Qiantang District, Hangzhou, Zhejiang Province